


NitrileCare Examination Gloves

Specifications & Certifications



Multi-Purpose



Product Technical Data Sheet

Type	Powder-Free, Examination Glove		
Specification	Non-Sterile/Disposable		
Cuff	Beaded		
Weight	Average 3.5g (0.12 fl oz) +/-0.3g		
Colour	Indigo Blue		
Primary Material	Nitrile		
Surface	External: Finger textured		
Powder Control	<=2mg/glove		
Packaging	100 pcs/box		
Size & Product Code	Small:	NG-NTCR-001	
	Medium:	NG-NTCR-002	
	Large:	NG-NTCR-003	
	Extra-large:	NG-NTCR-004	
Product Name	Nitrilecare Examination Gloves		
Brand	Careplus		
Application Setting	Medium risk medical, dental, procedures, pathology labs, patient care and food handling.		

Product Properties

Test Method	Characteristics			Requirement	Median
ASTM D6319	Dimensions (mm)	Width (M)		95 ± 10	95
		Length (M)		Min 230	240 (9.5")
		Thickness single (mm)	Finger	Min 0.05	0.10 (4.0mils)
			Palm	Min 0.05	0.08 (3.1mils)
ASTM D5151	Freedom from holes			No leakage	NA
ASTM D412	Before accelerated aging.				
	Tensile Strength (MPa)			Min 14	30
	Ultimate elongation (%)			Min 500	650
ASTM D573	After accelerated aging at (70±2 °C 166±2 hr).				
	Tensile Strength (MPa)			Min 14	14
	Ultimate elongation (%)			Min 400	400

Palm Width	Small: 80mm +/- 10	Weight	3.5g +/- 0.3g (0.12 fl oz)
	Medium: 95mm +/- 10	Length	9.5 inch glove
	Large: 110mm +/- 10	Thickness	3.5 mil glove
	Extra-large: 120mm +/- 10		

Certifications



US Standards	ASTM D6319, ASTM F1671 (Viral)
FDA Information	FDA 510(k) - K172015 Click here for FDA Website
Quality Standards	ISO 9001:2015, EN ISO 13485:2016,