



CENTRALBIDDING
FROM CENTRAL AUCTION HOUSE

M1491 Fire Department Automatic External Defibrillators
City of Baton Rouge - Parish of East Baton Rouge

Project documents obtained from www.CentralBidding.com

04-Mar-2024 09:33:14 AM



Office of the Mayor-President

Purchasing Division
City of Baton Rouge
Parish of East Baton Rouge
222 St. Louis St. 8th Floor Room 826
P.O.Box 1471

Baton Rouge, Louisiana 70821
225/389-3259 FAX 225/389-4841
purchasinginfo@brgov.com

Paul Narcisse
Director of Purchasing

ADDENDUM NO. 1
February 15, 2024

Your reference is directed to: **File Number: 24-01491**

Solicitation Number: **M1491 Fire Department Automatic External Defibrillators**

scheduled to open: **March 7, 2024**

The following changes will be made part of the above referenced solicitation.

This addendum is hereby officially made a part of the referenced solicitation and should be attached to the bidder's proposal or otherwise acknowledged therein.

If you have already submitted your proposal and this addendum causes you to revise your original bid, please indicate changes herein and return to Purchasing prior to bid opening in an envelope marked with the file number, bid opening date, and time. If this addendum does not cause you to revise your bid, please acknowledge receipt of the addendum by signing your name and company below and returning it in accordance with the provisions above.

cc: Bid File 24-01491

sharris@brla.gov
225-389-3259 x 3271

Signature

03/04/2024

Date

Uncharted Territory LLC

Company

ITEM 0001 Automated External Defibrillators (AED's)

The AED's provided by Contractor are required to:

- Device must be NEW.
- Device must be compatible with local EMS for pads and cables.
- Contractor must provide education and training.
- Contractor must supply newest software updates free of charge when available.
- Quickly capture and record patient ECG and heartrate.
- Contractor must supply loaner devices when repair is needed.
- Must have CPR Feedback device in AED.
- Have flexible operating modes to permit manual override to access more advanced capabilities, including heart rate displays, and to decide if and when to deliver treatment.
- Have a three-lead cable for basic monitoring using ECG electrodes.
- Have the capability to display the ECG trace within specific areas, always displaying gain to show the patient's rhythm.
- Capture, store, and have the ability to easily transfer rescue performance data, including, but not limited to, data on CPR rate, depth, release velocity, and compression fraction for CQI purposes.
- Be sufficiently light weight to reduce injury and improve working conditions among providers, The Contractor shall provide the overall dimensions and weight, inclusive of all peripherals, defibrillator cables, batteries and supplied option accessories in protective carrying case of the proposed unit.

The AED's shall include the following accessories:

- Protective carrying case (1) per device
- Adult/Pediatric Defibrillation Electrode (1) per device
- Pediatric Defibrillation Electrode (1) per device
- The AED's shall include a five (5) year parts and labor warranty against manufacturing defects and workmanship.
- The Contractor should also provide parts and labor covered under the warranty, exclusions of warranty, availability of warranty service agent(s), turnaround time on warranty repairs, availability of loaner AED, service options available after the warranty expires and authorized service agents.

ITEM 0002 BATTERY (Model: ZOLL AED Pro non-rechargeable lithium battery packs 8000-0880-01 or Approved Equal per Specifications)

Changed to

ITEM 0002 BATTERY (Model: ZOLL AED Pro non-rechargeable lithium battery packs 8000-0860-01 or Approved Equal per Specifications)

ITEM 0005 SOFTWARE SUBSCRIPTION (Model: ZOLL Case Review Premium 9400-110145 or Approved Equal per Specifications)

Changed to

ITEM 0005 SOFTWARE SUBSCRIPTION (Model: ZOLL Case Review Premium 8400-110145 or Approved Equal per Specifications)

PRICING

The brand and model used in this specification is used only to denote the general style, type, character and quality standard of the supplies requested. Please note the brand and model that you are bidding in the designated spaces provided. If bidding other than specified, sufficient information should be enclosed with the bid to in order to determine quality, suitability, compatibility and compliance with the specifications.

ITEM	DESCRIPTION	QUANTITY	UNIT	UNIT PRICE	Extended Price
0001	DEFIBRILLATOR (Model: ZOLL AED Pro 90110200499991010 or Approved Equal per Specifications) State Brand and Model Bid: ZOLL AED Pro	40	EACH	\$ 3250	\$ 130,000
0002	BATTERY (Model: ZOLL AED Pro non-rechargeable lithium battery packs 8000-0860-01 or Approved Equal per Specifications) State Brand and Model Bid: ZOLL Non rechargeable battery	40	EACH	\$ 215	\$ 8,600
0003	CABLE (Model: ZOLL ECG cable AAMI 8000-0838 or Approved Equal per Specifications) State Brand and Model Bid: ZOLL ECG Cable	40	EACH	\$ 190	\$ 7,600
0004	PADS (Model: ZOLL CPR Stat-padz HVP Multifunction CPR Electrodes [8 pair per case] 8900-0400 or Approved Equal per Specifications) State Brand and Model Bid: ZOLL CPR PADZ	10	CASE	\$ 720	\$ 7,200
0005	SOFTWARE SUBSCRIPTION (Model: ZOLL Case Review Premium 8400-110145 or Approved Equal per Specifications) State Brand and Model Bid: ZOLL Software subscription annual	2	EACH	\$ 395	\$ 790