



广州市哈雷日用品有限公司
Guangzhou Harley Daily Necessities Co., Ltd.

公司资料 (Company information)

编号: S2212019051300G(1-1)

统一社会信用代码

91440114076502373X

营业执照

(副本)

扫描二维码登录“国家企业信用信息公示系统”了解更多登记、备案、许可、监管信息。

名称

广州市哈雷日用品有限公司

注册资本

叁佰万元 (人民币)

类型

有限责任公司(自然人投资或控股)

成立日期

2013年08月15日

法定代表人

徐美英

营业期限

2013年08月15日 至 长期

经营范围

其他制造业 (具体经营项目请登录广州市商事主体信息公示平台查询, 网址: <http://cri.gz.gov.cn/>。依法须经批准的项目, 经相关部门批准后方可开展经营活动。)

住所

广州市从化区太平镇元洲岗村 (厂房A01)

登记机关

2019年 03月 21日

国家企业信用信息公示系统网址: <http://www.gsxt.gov.cn>

市场主体应当于每年1月1日至6月30日通过国家企业信用信息公示系统报送公示年度报告

国家市场监督管理总局监制

528361

对外贸易经营者备案登记表

备案登记表编号: 03629647 进出口企业代码: 4401076502373

经营者中文名称	广州市哈雷日用品有限公司		
经营者英文名称	Guangzhou Harley Commodity Company Limited		
组织机构代码	076502373	经营者类型 (由备案登记机关填写)	私营有限责任公司
住所	广州市从化区太平镇元洲岗村 (厂房A01)		
经营场所 (中文)	广州市从化区太平镇元洲岗村 (厂房A01)		
经营场所 (英文)	(PLANT A01) YUANZHOU GANG VILLAGE, TAIPING TOWN, CONGHUA DISTRICT, GUANGZHOU		
联系电话	020-37512951	联系传真	020-37512951
邮政编码	510900	电子邮箱	139036088@qq.com
工商登记注册日期	2013-8-15	工商登记注册号	91440114076502373X

依法办理工商登记的企业还须填写以下内容

企业法定代表人姓名	徐美英	有效证件号	342423197511104505
注册资金	叁佰万元		(折美元)

依法办理工商登记的外国 (地区) 企业或个体工商户 (独资经营者) 还须填写以下内容

企业法定代表人 / 个体工商户负责人姓名		有效证件号	
企业资产 / 个人财产			(折美元)

备注

填表前请认真阅读背面的条款, 并由企业法定代表人或个体工商户负责人签字、盖章。

备案登记机关

2019年 04月 10日

ISO 9001:2015 certificate:


认证证书

标准 ISO 9001:2015
证书登记号码 01 100 1430200
证书持有者: 广州市哈雷日用品有限公司
统一社会信用代码: 91440114076502373X
注册地址: 中国广东省广州市从化区太平镇元洲岗村 (厂房 A01)
邮编: 510900
经营地址: 同上述地址

认证范围: 口罩的设计和制造

证明完成了审核并满足了 ISO 9001:2015 标准的要求。
有效期: 证书有效期从 2019-09-08 至 2020-08-31。
此证书须经过符合要求的监督审核保持有效。
初次发证始于 2014 年
本证书信息可在国家认证认可监督管理委员会官方网站上查询
<http://www.cnca.gov.cn>

2019-09-10


TÜV Rheinland Cert GmbH
Am Grauen Stein · 51105 Köln

www.tuv.com



扫描全能王 创建

Certificate

Standard ISO 9001:2015
Certificate Registr. No. 01 100 1430200
Certificate Holder: Guangzhou Harley Commodity Company Limited
Unified Social Credit Code: 91440114076502373X
Registration Address: (Plant A01) Yuanzhougang Village,
Taiping Town, Conghua District, Guangzhou City,
Guangdong Province 510900, P. R. China
Operation Address: same as above

Scope: Design and Manufacturing of Respirator

Proof has been furnished by means of an audit that the requirements of ISO 9001:2015 are met.

Validity: The certificate is valid from 2019-09-08 until 2020-08-31.
It remains valid subject to satisfactory surveillance audits.
First certification 2014

This certificate information can be searched on CNCA official website <http://www.cnca.gov.cn>

2019-09-10


TÜV Rheinland Cert GmbH
Am Grauen Stein · 51105 Köln

www.tuv.com



扫描全能王 创建

Environmental management system certificate:

	
ENVIRONMENTAL MANAGEMENT SYSTEM CERTIFICATE	
Registration No. 0350319E20131R1S	
This is to certify that the environmental management system of GUANGZHOU HARLEY COMMODITY CO., LTD. Taiping Town, Yuanzhougang Village, Shunhua District, Guangzhou City (Factory Building A01), 510900 Social Credit Code: 91440114076502373X	
is in conformity with GB/T 24001-2016 / ISO 14001:2015 Standard	
This certificate is valid to Design and manufacturer of respirator (Site Covered: Taiping Town, Yuanzhougang Village, Shunhua District, Guangzhou City (Factory Building A01)).	
Date of issue: June 06, 2019 Date of expiry at most: April 26, 2022	 XINGYUAN CERTIFICATION CENTRE CO., LTD. (XQCC) FOR CERTIFICATE (7FL, Tower C, Jiahua Plaza, No. 9 Shangdi 3 St., Haidian, Beijing)
Representative: <i>Wang Honglin</i>	
	
中国认可 国际互认 管理体系 MANAGEMENT SYSTEM CNAS C035-M	
NOTE: This certificate shall be maintained by regular surveillance audit. The validity of the certificate can be verified by scanning QR code. The information of the certificate can be available in http://www.cnca.gov.cn , the website of CNCA, and in our website http://www.xqcc.com.cn .	
	

	
环境管理体系认证证书	
注册号: 0350319E20131R1S	
兹 证 明 广州市哈雷日用品有限公司 广州市从化区太平镇元洲岗村 (厂房A01), 510900 统一社会信用代码: 91440114076502373X	
管理体系符合 GB/T 24001-2016 / ISO 14001:2015 标准	
该管理体系适用于 口罩的设计和制造*** [涉及场所: 广州市从化区太平镇元洲岗村 (厂房A01)]	
颁证日期: 2019年6月6日 有效期最长可至: 2022年4月26日 ^注	 兴原认证中心有限公司 证书专用章 (北京市海淀区上地三街10号中农大厦C座7层)
签发人: <i>王启林</i>	
	
中国认可 国际互认 管理体系 MANAGEMENT SYSTEM CNAS C035-M	
注: 在证书有效期内, 获证组织须按规定接受年度监督审核, 保持认证资格。通过扫描二维码可获知 证书的有效状态。该证书信息还可在国家认证认可监督管理委员会官方网站 (www.cnca.gov.cn) 和兴原认证中心有限公司官方网站 (www.xqcc.com.cn) 上查询。	
	

Occupational health and safety management system certificate:


OCCUPATIONAL HEALTH AND SAFETY MANAGEMENT SYSTEM CERTIFICATE
Registration No. 0350319S30114R1S
This is to certify that the occupational health and safety management system of GUANGZHOU HARLEY COMMODITY CO., LTD. Taiping Town, Yuanzhougang Village, Shunhua District, Guangzhou City (Factory Building A01), 510800 Social Credit Code: 91440114076502373X
is in conformity with ISO 45001:2018 Standard
This certificate is valid to Design and manufacturer of respirator (Site Covered: Taiping Town, Yuanzhougang Village, Shunhua District, Guangzhou City (Factory Building A01)).
Date of issue: June 06, 2019 Date of expiry at most: April 26, 2022
Representative: <i>Wang Honglin</i>
 XINGYUAN CERTIFICATION CENTRE CO., LTD. (XQCC) (7FL Tower C, Jiahua Plaza, No. 9 Shangdi 3 St., Haidian, Beijing)
 中国认可 管理体系 MANAGEMENT SYSTEM CNAS C035-M
NOTE: This certificate shall be maintained by regular surveillance audit. The validity of the certificate can be verified by scanning QR code. The information of the certificate can be available in http://www.cnca.gov.cn , the website of CNCA, and in our website http://www.xqcc.com.cn .



职业健康安全管理体系认证证书
注册号: 0350319S30114R1S
兹 证 明 广州市哈雷日用品有限公司 广州市从化区太平镇元洲岗村 (厂房A01), 510800 统一社会信用代码: 91440114076502373X
管理体系符合 ISO 45001:2018 标准
该管理体系适用于 口罩的设计和制造*** [涉及场所: 广州市从化区太平镇元洲岗村 (厂房A01)]
颁证日期: 2019年6月6日 有效期最长可至: 2022年4月26日 ^注
签 发 人: <i>王宏林</i>
 兴原认证中心有限公司 证书专用章 (北京市海淀区上地三街9号嘉华大厦C座7层)
 中国认可 管理体系 MANAGEMENT SYSTEM CNAS C035-M
注: 在证书有效期内, 获证组织须按规定接受年度监督审核, 保持认证资格。通过扫描二维码可获知 证书的有效状态。该证书信息还可在国家认证认可监督管理委员会官方网站 (www.cnca.gov.cn) 和兴原认证中心有限公司官方网站 (www.xqcc.com.cn) 上查询。


产品介绍

Product introduction

N95 Respirator

L-188

N95 L188 mask picture:



NIOSH N95 (L-188) certificate:TN-19824



DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-19824

Mfr. Reference: GHCL-188

Centers for Disease Control
and Prevention (CDC)

National Institute for Occupational
Safety and Health (NIOSH)

National Personal Protective
Technology Laboratory (NPPTL)

P.O. Box 18070
Pittsburgh, PA 15236-0070

Phone: 412-386-4000

Fax: 412-386-4051

December 15, 2014

Mr. Wanghua Xu
Management Representative
Guangzhou Harley Commodity Company Limited
Floor 1 to 4, #8, Team 17, Sandong Village
Xinhua Street, Huadu District
Guangzhou 510800
CHINA

Dear Mr. Xu:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted September 17, 2014. This request was for approval of the model L-188 N95 filtering facepiece air purifying respirator for protections against particulates at a N95 filter efficiency level, reference assembly matrix L-188 Ama.xls. In addition, the Guangzhou Harley Commodity Company QA Manual, version A1, effective date 2014-06-10, was submitted for review.

This request is granted. Approvals are granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-84A-6973 has been assigned. This respirator is approved for protection against particulates at a N95 filter efficiency level.

NIOSH has also reviewed the quality manual presented and finds that this manual meets or exceeds the minimum technical requirements for quality assurance plans as outlined in Title 42, *Code of Federal Regulations* (CFR), Part 84.41(a) and, on the basis of this review, this quality manual is accepted.

The CD enclosed with this letter contains the final respirator label. The abbreviated label is accepted as submitted. The cautions and limitations which apply to this approval are on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

Page 2- Mr. Wanghua Xu - TN-19824

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

A copy of the quality manual will be retained by NIOSH and incorporated into our files. Any future changes to this accepted quality document must be submitted to NIOSH for a modification of this accepted quality system.

Sincerely yours,

Jonathan V. Szalajda
Acting Chief, Technology Evaluation Branch
National Personal Protective Technology Laboratory

Enclosures

NIOSH N95 (L-188) test report:TN-19824



TEST REPORT

Task Number: TN-19824
Manufacturer: Guangzhou Harley Commodity Company Limited
Prepared by: Nichole Petitta
Tests Conducted by: Nichole Petitta
Date: November 13, 2014
Respirator Tested: L-188 N95

Background Information

Respirator tested as per test request.

Tests Assigned

Test Description	STP Number
A. Exhalation Resistance Test	RCT-APR-STP-0003
B. Inhalation Resistance Test	TEB-APR-STP-0007
C. Sodium Chloride (NACL) N95 Test	TEB-APR-STP-0059

Overall Results

The respirator system tested met the requirements of all the procedures listed above.

Task Number: TN-19824 **Reference No.:** CFR 84.181
Test: Sodium Chloride (NaCl) - N95 **STP No.:** 59
Manufacturer: Guangzhou Harley Commodity Company Limited
Item Tested: L-188 N95

Comments:

All samoles were tested on TSI 8130 machine 000192 with samples 1-3 using timer 000214.

Was all equipment verified to be in calibration throughout all testing? ☒ Yes ☐ No

Signature: *Nichole L. Petitta* **Date:** 11/13/2014
Engineering Technician

NIOSH N95 (L-188) test report:TN-19824)

Task Number: TN-19824 Reference No.: CFR 84.180
Test: Exhalation Resistance Test STP No.: 3
Manufacturer: Guangzhou Harley Commodity Company Limited

Filter Type: Filter Only
Item Tested: L-188 N95

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Exhalation	Exhalation	
1	25	9.6	PASS
2	25	8.9	PASS
3	25	8.5	PASS

Overall Result: PASS

Comments:
Samples were tested on Brooks meter 0000283.
Was all equipment verified to be in calibration throughout all testing? ☒ Yes ☐ No

Signature:  Date: 11/4/2014
Engineering Technician

Task Number: TN-19824 Reference No.: CFR 84.180
Test: Inhalation Resistance Test STP No.: 7
Manufacturer: Guangzhou Harley Commodity Company Limited
Item Tested: L-188 N95

Filter Type: Filter Only

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Inhalation	Inhalation	
1	35	10.5	PASS
2	35	10.5	PASS
3	35	9.7	PASS

Overall Result: PASS

Signature:  Date: 11/4/2014
Engineering Technician

NIOSH N95 (L-188) test report:TN-19824

National Institute for Occupational Safety and Health
Respirator Branch
Test Data Sheet



Task Number: TN-19824 Reference No.: CFR 84.181
Test: Sodium Chloride (NaCl) - N95 STP No.: 59
Manufacturer: Guangzhou Harley Commodity Company Limited
Item Tested: L-188 N95

Filter	Flow Rate	Initial Filter Resistance	Maximum Allowable Percent Leakage	Initial Percent Leakage	Maximum Percent Leakage	Result
1	85	9.6	5.00	0.517	0.517	PASS
2	85	10.6	5.00	0.694	0.694	PASS
3	85	9.8	5.00	0.718	0.718	PASS
4	85	10.3	5.00	0.651	0.654	PASS
5	85	9.3	5.00	0.682	0.682	PASS
6	85	9.7	5.00	0.755	0.755	PASS
7	85	10.7	5.00	0.621	0.621	PASS
8	85	9.9	5.00	0.744	0.744	PASS
9	85	9.8	5.00	0.535	0.535	PASS
10	85	10.8	5.00	1.500	1.580	PASS
11	85	10.5	5.00	0.688	0.688	PASS
12	85	9.7	5.00	0.692	0.692	PASS
13	85	10.6	5.00	0.757	0.757	PASS
14	85	10.8	5.00	0.671	0.671	PASS
15	85	10.8	5.00	0.981	0.981	PASS
16	85	10.3	5.00	0.678	0.678	PASS
17	85	9.8	5.00	0.661	0.661	PASS
18	85	10.3	5.00	0.680	0.680	PASS
19	85	9.6	5.00	1.660	1.660	PASS
20	85	9.8	5.00	0.769	0.769	PASS

Overall Result: PASS

Task Number: TN-19824 Reference No.: CFR 84.180
Test: Inhalation Resistance Test STP No.: 7
Manufacturer: Guangzhou Harley Commodity Company Limited
Item Tested: L-188 N95

Comments:
Samples were tested on Brooks meter 000036.

Was all equipment verified to be in calibration throughout all testing? ☒ Yes ☐ No




Signature: *Nicole L. Petitta* Date: 11/13/2014
Engineering Technician

Signature: *Nicole L. Petitta* Date: 11/4/2014
Engineering Technician

L-188 实物图片 Pictures



L-188 包装盒图片 Packing Pictures



盒规 (Box) : 20Pcs/Box,Size:200*170*135

L-188 合格证、箱规 Certificate &CTN Details



箱规 (CTN) : 20Boxes/CTN Size:415*415*345 G.W:4.9KG