



With ultimate hygiene in mind, Cranberry products are First Touch® manufactured, examined, and packaged with zero direct skin contact exposure. Look for our First Touch® logo and be assured that you are doing the best you can to protect yourself and your patients.

The Evolve 300 Nitrile Powder Free Examination Gloves are Cranberry's latest and softest gloves yet, combining comfort and tensile strength without sacrificing tactile sensitivity. Cranberry's exclusive EvoSoft™ formulation gives Evolve a unique silk-like attribute that is both soft and strong with textured fingertips for precise gripping under all operating conditions.

300
evolve®
Nitrile Powder Free
Examination Gloves

Cranberry Evolve Ultrasoft Powder Free Nitrile Gloves

Exclusive EVOSOFT™ formulation offers unprecedented softness and comfort.

- ✓ Superior tensile strength for extended wear
- ✓ Nitrile Powder Free
- ✓ Latex Free
- ✓ Non sterile
- ✓ Dark purple colour
- ✓ Single use only
- ✓ Fits glove dispenser code: AC064





FORMFITTING

EvoSoft™ formulation delivers a silk-like attribute, soft and strong. Fingertip textured in a Royal Blue color.

SILK THIN

Our scratchiest gloves yet, high tensile strength and excellent comfort, while not sacrificing tactile sensitivity.



300 GLOVES PER BOX

Ultra 300 Saver Pack reduces storage space and packaging waste.





evolve³⁰⁰™ Nitrile Powder Free Exam Gloves

Features	Benefits
Exclusive EvoSoft™ formulation	Provides unprecedented soft and strong comfort during extended wear
Textured Fingertips	Delivers good grip
300 Space Saver Pack	Maximize storage space and efficiency

Recommended for use when:

- i) A latex-free, powder-free solution is required.
- ii) Seeking a soft and tear-resistant, protective solution with good grip.
- iii) Seeking space-saving, reduced waste packaging.



Physical Data					
Dimension	Extra-Small	Small	Medium	Large	Extra-Large
Length (mm) min.	230	230	230	230	230
Palm Width (mm)	70±10	80±10	95±10	110±10	120±10
Thickness (mm) min.					
- Finger	0.05	0.05	0.05	0.05	0.05
- Palm	0.05	0.05	0.05	0.05	0.05
- Cuff	0.04	0.04	0.04	0.04	0.04
Tensile Strength (MPa), min					
- Unaged	16	16	16	16	16
- Aged	16	16	16	16	16
Elongation (%), min					
- Unaged	500	500	500	500	500
- Aged	400	400	400	400	400

Single Use Only, Ambidextrous, Non-Sterile

Packaging: 300 gloves/box, 10 boxes/case; XL: 250 gloves/box

XS (3305) **M** (3307) **XL** (3309)

S (3306) **L** (3308)

 www.cranberryusa.com
like us on  facebook.com/cranberryusa

Cranberry USA
828 San Pablo Avenue
Pinole, CA 94564



Passionate about your health

Cranberry world leader in Quality & range of Gloves for your total protection

Stop buying cheap, rejected, repacked gloves with inconsistent qty, self-branded, without Certificate.

Your safety is our concern.
Confident brings world's top class, award winning brand with all standard Certificates.

Verify yourself, decide your purchase.

AWARDS



CERTIFICATIONS



DRIVING BY INNOVATION



Specifications



- Cranberry Evolve Powder Free Nitrile Disposable Gloves Box of 300 / 150 Pairs
- Product Type High tensile strength Latex and Powder Free Nitrile Disposable Ambidextrous Gloves
- Textured fingertips that feel like the touch of skin!
- Nitrile gloves material is the way-of-the-future for all medical-grade/surgical gloves.
- ODOURLESS! Will not leave your hands with an unpleasant odor!
- NON-SLIPPERY! No slipping, even when wet!
- Sold as a Box 300 Gloves / 150 pairs
- Design and packaging may vary from batch to batch.
- Color Dark Blue



Personal protective equipment is a cornerstone of modern infection control protocol, and Cranberry USA is committed to equipping oral health professionals with the tools they need to deliver safe and effective care. Evolve nitrile powder-free examination gloves represent the company's softest yet most durable gloves to date. Packaged in a 300-glove-count box, the product not only provides protection but also reduces packaging waste and enhances cost savings.

TACTILE SENSITIVITY

Designed to precisely contour to practitioners' hands, Evolve gloves offer a custom fit and feel while delivering protection and lightweight comfort. Made with the company's proprietary EvoSoft formulation, which is designed to maximize protection, durability, and chemical resistance, Evolve exam gloves are one of the thinnest on the market and provide excellent tactile sensitivity. Textured fingertips offer superior dexterity, ensuring a secure grip in both wet and dry conditions.

Evolve is engineered with Cranberry USA's pioneering First Touch® technology, which ensures no previous skin contact exposure during manufacturing and delivery to dental offices. They're available in sizes XS through XL.

Test Report No. 7191
dated 23 Mar 2018



PSB Singapore

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

Testing of Powder Free Nitrile Examination Gloves submitted by
Co., Ltd. on 25 Jan 2018 and 09 Mar 2018.

Choose certainty.
Add value.

TESTED FOR:

Shandong,
China.

TEST DATE:

26 Jan 2018 to 06 Feb 2018 and 22 Mar 2018

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Reference No.	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Powder Free Nitrile Examination Gloves	Blue	BS0002	12150511	XS	100	Ltd.
				03060511		69	
				12150521	S	100	
				12080311	M	100	
				12090411	L	100	
				12070611	XL	407	

Lot size as specified by client: 200,000 pieces

METHOD OF TEST:

- EN 455-1:2000 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes
- EN 455-2:2015 Medical gloves for single use
Part 2: Requirements and testing for physical properties
- EN 455-3:2015 Medical glove for single use
Part 3: Requirements and testing for biological evaluation



Laboratory:
TUV SUD PSB Pte. Ltd.
No. 1 Science Park Drive
Singapore 118221

www.tuv-sud-psb.sg
Co. Reg : 199002667R

Regional Head Office:
TUV SUD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221

Test Report No. 719
dated 23 Mar 2018



PSB Singapore

RESULTS:

Sample: Powder Free Nitrile Examination Gloves, BS0002

Table 1: Results for EN 455-1:2000

Clause	Tests	Size	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4 5	Freedom from holes	XS	Shall not leak	10	315	1	Passed
		S		10	315	1	Passed
		M		10	315	2	Passed
		L		10	315	6	Passed
		XL		10	315	4	Passed

Table 2: Results for EN 455-2:2015 Clauses 4-5

Clause	Tests	Size	Requirements (Median)	Number tested (pieces)	Results (Median)	Inferred results
4	Dimensions a) Length (mm)	XS	≥ 240	13	250	Passed
		S		13	250	Passed
		M		13	255	Passed
		L		13	250	Passed
		XL		13	248	Passed
	b) Width (mm)	XS	≤ 80	13	73	Passed
		S	80 ± 10	13	85	Passed
		M	95 ± 10	13	96	Passed
		L	110 ± 10	13	106	Passed
		XL	≥ 110	13	115	Passed
5	Strength a) Force at break (N)	XS	For nitrile examination gloves: ≥ 6.0	13	6.4	Passed
		S		13	8.6	Passed
		M		13	6.1	Passed
		L		13	6.1	Passed
		XL		13	6.6	Passed
	b) Force at break after challenge testing (N)	XS	For nitrile examination gloves: ≥ 6.0	13	7.0	Passed
		S		13	9.0	Passed
		M		13	7.3	Passed
		L		13	6.6	Passed
		XL		13	7.1	Passed

Table 3: Results for EN 455-2:2015 Clause 7

Clause	Tests	Requirements	Results	Inferred results
7	Labelling	Manufacturers shall label the glove and/or the packaging with the date of manufacture in accordance with EN ISO 15223-1:2012 and EN 1041:2008+A1:2013. Date of manufacture is defined as the packaging date.	Observed	Passed

Test Report No. 7191
dated 23 Mar 2018



PSB Singapore

RESULTS (cont'd):

Sample: Powder Free Nitrile Examination Gloves, BS0002

Table 4: Results for EN 455-3:2015 Clauses 4.2-4.5

Clause	Tests	Requirements	Results / Remarks	Inferred results
4.2	Chemicals	Gloves shall not be dressed with talcum powder (magnesium silicate).	Glove is powder-free glove, based on client's declaration letter version 2018001	NA
		Other chemicals	Manufacturer shall disclose upon request a list of chemical ingredients	NA
4.3 5.1	Endotoxins	< 20 EU/pair for gloves labelled with 'low endotoxin content'.	Not labelled with 'low endotoxin content'	NA
4.4 5.2	Powder-free gloves	For powder-free gloves: The total quantity of powder residues shall not exceed 2 mg per glove.	XS 0.61 mg per glove	Passed
			S 0.63 mg per glove	Passed
			M 0.97 mg per glove	Passed
			L 0.86 mg per glove	Passed
			XL 0.48 mg per glove	Passed
4.5 5.3	Proteins, leachable	The manufacturer shall strive to minimize the leachable protein level for gloves containing natural rubber latex.	Non-natural rubber latex glove	NA

Table 5: Results for EN 455-3:2015 Clause 4.6

Clause	Tests	Requirements	Results
4.6	Labelling	In addition to the labelling specified in EN 1041:2008+A1:2013 and the relevant symbols given in EN ISO 15223-1:2012, the following requirements apply:	
		a) medical gloves containing natural rubber latex shall be labelled on the packaging of at least the smallest packaging unit with the EN ISO 15223-1:2012 symbol for latex;	NA
		The labelling shall include the following or equivalent warning statement together with the symbol: '(Product) contains natural rubber latex which may cause allergic reactions, including anaphylactic responses';	NA
		b) the labelling shall include a prominent indication of whether the glove is powdered or powder-free;	Comply
		c) sterile powdered gloves shall be labelled with the following or equivalent: 'CAUTION: Surface powder shall be removed aseptically prior to undertaking operative procedures in order to minimize the risk of adverse tissue reactions';	NA
		d) for any medical glove containing natural rubber latex the product labelling shall not include: - any term suggesting relative safety, such as low allergenicity, hypoallergenicity or low protein; - any unjustified indication of the presence of allergens;	NA
		e) if the manufacturer labels the gloves with the protein content, the process limit, measured as specified in 5.3 shall be given.	NA
Inferred results			Passed

Test Report No. 7191 [REDACTED]
dated 23 Mar 2018



REMARKS:

1. Freedom from holes test for XS, S, M and L sizes were tested in manufacturer's site, witnessed by TÜV SÜD Certification and Testing (China) Co., Ltd. Beijing Branch on 22 Mar 2018.
2. For size XS, results for EN 455-2:2015 Clause 4 Dimensions is based on lot no. 03060511, while the rest of the results are based on lot no. 12150511.
3. Labelling requirements are assessed based on submitted packaging artwork together with client's declaration letter version number 2018003.
4. NA: Not applicable for the submitted sample.


Shareen Chan
Engineer


Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX:



Photo : Powder Free Nitrile Examination Gloves, BS0002



U.S. Department of Health & Human Services

U.S. FOOD & DRUG ADMINISTRATION

Follow FDA | En Español

SEARCH

Home | Food | Drugs | Medical Devices | Radiation-Emitting Products | Vaccines, Blood & Biologics | Animal & Veterinary | Cosmetics | Tobacco Products

510(k) Premarket Notification

[FDA Home](#) | [Medical Devices](#) | [Databases](#)

510(k) | DeNovo | Registration & Listing | Adverse Events | Recalls | PMA | HDE | Classification | Standards
CFR Title 21 | Radiation-Emitting Products | X-Ray Assembler | Medsun Reports | CLIA | TPLC

[New Search](#) | [Back To Search Results](#)

Device Classification Name	Polymer Patient Examination Glove
510(K) Number	K983334
Device Name	NITRILE POWDER FREE EXAMINATION GLOVES
Applicant	CRANBERRY (M) SDN. BHD. LOT 85, JALAN PORTLAND, TASEK INDUSTRIAL ESTATE Ipoh Perak, MY 31400
Applicant Contact	Chong Yoon Tat
Correspondent	CRANBERRY (M) SDN. BHD. LOT 85, JALAN PORTLAND, TASEK INDUSTRIAL ESTATE Ipoh Perak, MY 31400
Correspondent Contact	Chong Yoon Tat
Regulation Number	880.6250
Classification Product Code	LZA
Date Received	09/23/1998
Decision Date	02/19/1999
Decision	Substantially Equivalent (SESE)
Regulation Medical Specialty	General Hospital
510k Review Panel	General Hospital
Statement	Statement
Type	Traditional
Reviewed By Third Party	No
Combination Product	No

Page Last Updated: 06/29/2020
Note: If you need help accessing information in different file formats, see Instructions for Downloading Viewers and Players.
Language Assistance Available: Español | 繁體中文 | Tiếng Việt | 한국어 | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | فارسی | English

Accessibility | Contact FDA | Careers | FDA Basics | FOIA | No FEAR Act | Nondiscrimination | Website Policies

U.S. Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
Ph. 1-888-INFO-FDA (1-888-463-6332)
Contact FDA

For Government | For Press

Combination Products
Advisory Committees
Science & Research
Regulatory Information
Safety
Emergency Preparedness
International Programs
News & Events
Training and Continuing Education
Inspections/Compliance
State & Local Officials
Consumers
Industry
Health Professionals

U.S. Department of Health & Human Services



CRANBERRY (M) SDN. BHD. (104994 - W)
(A MEMBER OF THE YEE LEE GROUP)

Applicant : CRANBERRY (M) SDN. BHD.

510(k) Number (if known): K*83*3*

Device Name: Nitrile Powder Free Examination Glove - Cranberry

Indications For Use:

This product is a patient examination glove. It is a disposable device intended for medical purposes that is worn on the examiner's hand or finger to prevent contamination between patient and examiner.

(PLEASE DO NOT WRITE BELOW THIS LINE - CONTINUE ON ANOTHER PAGE
IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Chin S. Lim
(Division Sign-Off)
Division of Dental, Infection Control,
and General Hospital Devices
510(k) N K98****

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter-Use X



Cranberry Nitrile Gloves Standard Operating Procedure

1. Sign NCNDA agreements followed by ICPO and POF
2. SELLER to respond to ICPO with models, pricing and quantity. (Note: models available will be determined on a case by case basis)
3. Draft SLA/SPA-conditions to be reviewed by BUYER. If totally in agreement, continue to the next stage.
4. The BUYER will send an ICPO to the SELLER or to a company as directed by the SELLER (Numeric Advisory). BUYER to specify exact model, pricing and quantities of each size in the ICPO
5. The SELLER will submit a pro-forma invoice to the BUYER.
6. Bank Guarantee/LC/Escrow must be set up to move to the next stage. (NOTE: No videos or proof of life will be provided in this process.)
7. The BUYER will accept the proforma invoice and both parties will sign a service level agreement (SLA/SPA).
8. The BUYER opens draft LC into 2 accounts (for factory and agent)
9. SGS certificates will be made available 48 to 72 hrs. after the order has been placed. The certificate will indicate that the stock will meet the regulated requirements. Delivery schedule is provided.
10. A waybill will be issued to the BUYER informing them that the stock is in transit.
11. Once the stock lands, the stock will be viewed by the BUYER and the SELLER and the remaining 90% of the funds will be released to the SELLER
12. The SELLER will transfer the SGS certificate to the name of the BUYER.

LC release by the BUYER (F.O.B Term)

13. The BUYER then owns the stock and is responsible for any fees and costs thereafter.
14. The BUYER's agents will clear this through customs.
15. We would highly recommend that the appropriate transportation, security, and insurance is arranged by the BUYER.
16. This is a CIF transaction.